

## CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_191122_1209
Address	F-319, S.I.T.E Karachi	Date of Issue	18-11-22
E mail	<a href="mailto:Mustafain.baig@searlecompany.com">Mustafain.baig@searlecompany.com</a>	Date of Expiry	17-05-23

Unit Under Test UUT			
User ID	AMP-HOTC-002	Sub Location	GRANULATION
Location	HOTC	Instrument Type	Ampere Meter
Model	AUTONICS	Make	AUTONICS
Range	0-500 AMP	Accuracy/Resolution	1 AMP
Serial #	N/A	Condition	PASS
Least Count	1 AMP	Tolerance Limit	± 1 AMP

Master / Testing / Reference Equipment Used			
Digital Multimeter			
Model	UNI-T/139S	Serial #	203604417
Range	0-600 Volt	Accuracy	0.01 Volt
Traceability Ref #	ILS/0322/178/01	Cal Due date	2023-03-04

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
AMP	AMP	AMP	AMP
0	0	0.00	PASS
100	100	0.00	PASS
200	200	0.00	PASS
300	300	0.00	PASS

\*UUT = Unit under Test \* N/A= Not applicable

**Remarks:** Above given Ampere Meter working as per manufacture specifications.



\_\_\_\_\_  
Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor  
ISO/IEC17025/APLACTR001*

\_\_\_\_\_  
Company Stamp

\_\_\_\_\_  
TheSearleCompany Representative