

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_191122_1204
Address	F-319, S.I.T.E Karachi	Date of Issue	21-11-22
E mail	Mustafain.baig@searlecompany.com	Date of Expiry	20-05-23

Unit Under Test UUT			
User ID	AM-SP-001	Sub Location	AMPOULE WASHING #02
Location	STERILE PRODUCTION	Instrument Type	Ampere Meter
Model	N/A	Make	N/A
Range	0-75 AMP	Accuracy/Resolution	1 FS
Serial #	N/A	Condition	PASS
Least Count	1 FS	Tolerance Limit	± 1 FS

Master / Testing / Reference Equipment Used			
Digital Multimeter			
Model	UNI-T/139S	Serial #	203604417
Range	0-600 Volt	Accuracy	0.01 Volt
Traceability Ref #	ILS/0322/178/01	Cal Due date	2023-03-04

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
AMP	AMP	AMP	AMP
0	0	0.00	PASS
55	55	0.00	PASS
60	60	0.00	PASS
65	65	0.00	PASS

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Ampere Meter working as per manufacture specifications.



Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001*

Company Stamp

TheSearleCompany Representative