

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_211122_1459
Address	F-319, S.I.T.E Karachi	Date of Issue	04-10-22
E mail	Mustafain.baig@searlecompany.com	Date of Expiry	03-04-23

Unit Under Test UUT			
User ID	CM-CU-005	Sub Location	RO PLANT
Location	CENTRAL UTILITIES	Instrument Type	Conductivity Meter
Model	FWT	Make	FWT
Range	MULTIPLE μ s	Accuracy/Resolution	1 %
Serial #	N/A	Condition	PASS
Least Count	1 μ s	Tolerance Limit	\pm 1 μ s

Master / Testing / Reference Equipment Used			
Conductivity Solution			
Model	Extech	Serial #	000002
Range	1280 , 1413 μ s/cm	Accuracy	1 %
Traceability Ref #	EXTECH	Cal Due date	2022-12-30

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
μ s	μ s	μ s	μ s
0	0	0.00	PASS
1413	1413	0.00	PASS
1413	1413	0.00	PASS
1413	1413	0.00	PASS

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Conductivity Meter working as per manufacture specifications.



Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001*

Company Stamp

TheSearleCompany Representative