

## CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_211122_1416
Address	F-319, S.I.T.E Karachi	Date of Issue	11-10-22
E mail	<a href="mailto:Mustafain.baig@searlecompany.com">Mustafain.baig@searlecompany.com</a>	Date of Expiry	10-04-23

Unit Under Test UUT			
User ID	ANM-RMS-01	Sub Location	Main Dispensing
Location	Raw Material store	Instrument Type	Anemometer
Model	TES-1341	Make	China
Range	0.00-30.00 m/s	Accuracy/Resolution	0.00 m/s
Serial #	N/A	Condition	PASS
Least Count	0.1 m/s	Tolerance Limit	± 1 m/s

Master / Testing / Reference Equipment Used			
Anemometer			
Model	TA-440	Serial #	4401316005
Range	0.00-30.00 m/s	Accuracy	0.025 m/s
Traceability Ref #	4401316005	Cal Due date	2023-01-06

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
m/s	m/s	m/s	m/s
0.45	0.45	0.00	Pass
0.45	0.45	0.00	Pass
0.45	0.45	0.00	Pass

\*UUT = Unit under Test \* N/A= Not applicable

**Remarks:** Above given Anemometer working as per manufacture specifications.



\_\_\_\_\_  
Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor  
ISO/IEC17025/APLACTR001*

\_\_\_\_\_  
Company Stamp

\_\_\_\_\_  
TheSearleCompany Representative