

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_211122_1460
Address	F-319, S.I.T.E Karachi	Date of Issue	17-11-22
E mail	Mustafain.baig@searlecompany.com	Date of Expiry	10-04-23

Unit Under Test UUT			
User ID	CG-PL-010	Sub Location	AUTOCLAVE AREA
Location	PEDITRAL LIQUID	Instrument Type	Compound Gauge
Model	BLD	Make	BLD
Range	-0.1 ~ 0.5 Mpa	Accuracy/Resolution	0.02 Mpa
Serial #	N/A	Condition	PASS
Least Count	0.02 Mpa	Tolerance Limit	± 0.02 Mpa

Master / Testing / Reference Equipment Used			
Digital Clamp Meter			
Model	UT207B	Serial #	203561411
Range	0-1000 AMP	Accuracy	0.1 AMP
Traceability Ref #	ILS/0322/178/02	Cal Due date	2023-03-04
Hand Pressure Pump			
Model	PV-212	Serial #	N/A
Range	0-700 Bar	Accuracy	N/A Bar
Traceability Ref #	N/A	Cal Due date	N/A

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
Mpa	Mpa	Mpa	Mpa
0.1	0.1	0.00	PASS
0.2	0.2	0.00	PASS
0.3	0.3	0.00	PASS
0.4	0.4	0.00	PASS

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Compound Gauge working as per manufacture specifications.



Noman Khan (Test Eng.)

PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001

Company Stamp

TheSearleCompany Representative